


INSTRUCTION MANUAL LABVIEW DRIVER



For the AVS47-IB & AVS-47B & TS-530A

LabView version 7.1



RV-Elektroniikka Oy PICOWATT
Veromiehentie 14
FI-01510 VANTAA, Finland
phone +358 50 337 5192 or +358 9 822087
fax +358 9 822184
Internet: www.picowatt.fi



CONTENTS

1. INTRODUCTION	4
2. ABOUT THE DESIGN PHILOSOPHY	4
VIRTUAL INSTRUMENTS	5
3. THE FIRST AND THE LAST	5
3.1 Initialize	5
3.2 Close	6
4. CONFIGURATION VI'S	6
4.1 Configure AVS47	6
4.2 Configure AVS47IB	7
4.3 Configure addresses	8
4.4 Configure TS530A	9
4.5 Enable service requests	10
4.6 Configure time and date	10
4.7 Configure scan parameters	10
4.8 Configure channel parameters	11
4.9 Configure reference voltage	12
5. ACTION/STATUS VI'S	12
5.1 Go remote	12
5.2 Go local	13
5.3 Poll STB status byte	13
5.4 Start single ADC	14
5.5 Wait for MAV/OPC/TMO	15
5.6 Start averaging	15
5.7 Start digital filter	15
5.8 Stop averaging etc.	16
5.9 Start single scan	16
5.10 Start continuous scan	17
6. DATA VI'S	17
6.1 Read single ADC	17
6.2 Read one reading	18
6.3 Read average	19
6.4 Read digital filter	20
6.5 Read scanresults	21
6.6 Read and save the buffer	21
6.7 Read AVS47 parameters	22
6.8 Read from TS530A	23
6.9 Read TS530A parameters	23
6.10 Read AVS47IB parameters	24
6.11 Read channel parameters	25
6.12 Read scan parameters	25
7. UTILITY VI'S	26
7.1 Reset	26
7.2 Default setup	26
7.3 Self test	26
7.4 Self calibrate	26



8. OMITTED VIs	27
Revision Query.vi	27
Error Query.vi	27
Error Message.vi	27
9. APPLICATION EXAMPLES	28
9.1. Discuss with AVS47IB.vi	28
9.2. Getting started.vi	29
9.3. Application example 1.vi	30
9.4. Scan 2 ch.vi	33
INDEX	34
10. REVISION HISTORY	37



1. INTRODUCTION

This LabView driver for the AVS47-IB covers all the functionality of the AVS-47B Resistance Bridge, the TS-530A Temperature Controller and the intelligent AVS47-IB Secondary Interface Unit. The driver more or less conforms to National Instruments' recommendations (or "standards" as NI says), consisting of more than 40 VI's. The driver is released without any error handling. However, most of the virtual instruments have been written so that they are just bypassed if they cannot work properly, and the VI's themselves often provide error indicators which can be inspected by opening the VI.

Instead of writing drivers separately for the AVS-47B and the TS-530A, we have handled both instruments within the same package. This corresponds to the facts that these instruments are mostly used together, and that the TS-530A without a resistance bridge is of little value. The AVS47-IB interface box contains many powerful features like computer-independent averaging and scanning. One purpose of this driver is to enhance the use of these features.

This driver has been written using a minimum of VISA functions, so compatibility with future platforms should be good. You need not devote your finest computer for running our driver. The first version of this driver was tested on a 166MHz Windows 3.11 computer, using an old National Instruments GPIB-PCII card whereas the current version 1.2 was tested on a 1GHz Pentium computer running Windows XP. This version was compiled for LabView version 7.1.

NOTE: Check the version of your driver by opening the "AVS47IB VI-Tree.vi" and looking the "Show VI Info" and "Show History".

2. ABOUT THE DESIGN PHILOSOPHY

The AVS47-IB is not just a dummy protocol converter between Picobus and IEEE-488 interface standards. The fact that the secondary interface unit contains an embedded PC made it natural to incorporate many programmable features like recording averages or scanning several sensors at predetermined intervals. Also buffering of the results was included, as it can sometimes solve time-critical problems.

This kind of features can of course be implemented using an external computer. But if the computer is used to control and read a large number of laboratory instruments, then the workload can make the computer too slow for a timely response or for real-time data acquisition. Despite the development in computer performance and even if LabView is now much faster than years ago, you may encounter timing problems in a large system.


Some help to the performance problem can be achieved if any or some of the instruments can perform time-taking tasks without the computer's supervision. As far as the AVS47-IB is concerned, examples of such slow tasks are averaging and scanning with or without simultaneous temperature control.

According to National Instruments' recommendations, instrument drivers should contain an **Initialize.vi** that sets the instrument into a known initial state. The drivers should also contain a **Close.vi** that leaves the instrument in a known state. We have deviated from these recommendations by purpose. Our **Initialize.vi** sets the system in remote control mode, but it *does not change the state of the instruments*. Our **Close.vi** does even less - it leaves the system in local control mode, but you can also tell it to leave it in the remote mode. Contrary to NI's recommendations, we have also included a number of VI's for *reading the current state* of the system into the LabView. The purpose of all this was to

- allow LabView to use more of its capacity for other tasks
- allow the user to close some virtual instruments in order to make more memory free.
- allow the user to exit LabView and to use the computer for something else
- allow the user to shut off and later re-start the computer, without disturbing the system, and without disturbing the controlled temperature in the cryostat.

The AVS-47B uses AC excitation whose frequency is only 12.5 or 13.7Hz. Therefore the instrument is very slow, and the measurement commands need a long time to complete. This driver makes extensive use of both the *OPC (operation complete) command and serial polling of the IEEE-488.2 defined Status Byte (STB) in loops waiting either for the operation to complete or for a message to become available (MAV). These loops give their waiting time to other simultaneously running VI's.

Throughout this driver you can see one **common structure: the VISA read and write operations are enclosed in a CASE statement**. The case is TRUE if the AVS47-IB is in idle state. There are only a few commands that the interface can accept if it is not idling, and these situations are handled differently. Usually, the idle state is first verified, and the VISA operation is allowed after that. If the system was not idling, the VISA operation is not performed, and usually some error indicator (like "not in idle state") is set.



Following National Instruments' standards, communications with the GPIB controller card have been realized using VISA. We needed five VISA functions in this driver: VISA open, VISA Read STB, VISA Write, VISA Read, and VISA Close.

VIRTUAL INSTRUMENTS

3. THE FIRST AND THE LAST

3.1 Initialize



Within LabView, **Initialize.vi** establishes communications between the software, the GPIB card and the GPIB device. For that purpose it needs to know the bus address of the AVS47-IB. The factory default address is 20 (refer also to **Configure addresses.vi**). Further, the LabView application which uses this particular AVS47-IB in this particular address is allocated a unique handle, called "VISA session". All the subsequent VI's that belong to

this application are normally connected together using this VISA session handle. The VI's are like inserted in a thread, which controls the execution order of the VI's in a multitasking environment.

Within the AVS47-IB, **Initialize.vi** disables the front panel started measurements (the START/PRINT button on the AVS-47B front panel). Such measurements must not be made in the remote control mode. The command for disabling them FSM0.

The response message headers are also disabled. After power-on, the AVS47-IB includes headers in all responses. For example, response to the *STB? query is STB 16. After the headers have been disabled by HDR 0, the response to *STB? will be just the number 16. Responses that contain only numbers, possibly separated by semicolons, are much easier to use in LabView as they need no parsing.

The operation complete -event is enabled for status reporting by the common command *ESE1. Each time the 47IB encounters the *OPC command, it sets bit 2⁰ of the ESR status register. After the *ESE1 command, this bit will be OR'ed into bit 2⁵ of the Master Status Byte register STB. By means of polling the STB (serial polling) one can determine whether the operation or a longer series of operations has completed.

These enabling and disabling actions are performed in the **Default setup.vi**, inside the **Initialize.vi**

The system is set into remote control mode using REM1.

The **Initialize.vi** offers two important choices:

- 1) **Identification query.** If enabled, the AVS47-IB will receive the query *IDN?, and it will answer with response PICOWATT,AVS47-IB,0,3.3 (for revision 3.3) . If the AVS47-IB is not idling, do not enable this query. As default, the query is disabled. It is safest to keep it disabled, unless you necessarily need to verify the kind of the device or the version of the firmware.
- 2) **Reset.** The default is not to reset neither the AVS-47B nor the TS-530A, as this can lead to unwanted changes in the system state or controlled temperature. You can reprogram all aspects of the system without first resetting it. If you want to avoid changes in the system state, find the system status by first reading it. Refer to the AVS47-IB manual for the reset states of these instruments.

Using these defaults, you can initialize the VISA session without disturbing the system.

Note that the **Default setup.vi** does not run if the AVS47-IB is not idling. The state is verified by polling the STB.

3.2 Close



The **Close.vi** contains two actions. First, it runs the **Go Local.vi** that sets the AVS-47B in the local control mode using REM0. Here again, as with most of the VI's in this driver, the status byte is first polled in order to find whether the 47IB is idling. If it is not idling, the **Go Local.vi** is just bypassed without any error message.

As default, going to local is enabled. However, you may want to disable this command, so that the AVS-47B remains in remote mode. A reason to do this could be, for example, that you want to prevent anybody from changing the bridge settings. Disabling is done from the LV front panel of the **Close.vi** or by using the corresponding Boolean control.

As the last thing, **Close.vi** closes the VISA session.

If you want to stop computer control only momentarily, consider disabling the REM0 command. Then the system stays in the remote mode, the front panel switches of the AVS-47B are disabled and unintentional changing of the system state is impossible.

4. CONFIGURATION VI'S

According to their definition, configuration VI's change the state of the instrument, but they do not produce any data which should be read. Configuration VI's are used for setting various parameters prior to a measurement so that the measurement can be started by using a simple command.

Because this driver actually covers three devices, the AVS47-IB, the AVS-47B and the TS-530A, the list of configuration VI's is long.

4.1 Configure AVS47




Configure AVS47.vi covers the front panel switches of the AVS-47B:

- Input:** 0..2 (zero, measure, calibrate). The GPIB command to AVS47-IB is INPx.
- Channel:** 0..7. The AVS47-IB command is MUXn
- Range:** 0..7 (no range, 2Ω, 20Ω, 200Ω, 2kΩ, 20kΩ, 200kΩ, 2MΩ). The AVS47-IB command is RANx
- Excitation:** 0..7 (no exc, 3μV, 10μV, 30μV, 100μV, 300μV, 1mV). The AVS47-IB command is EXCn
- Display:** 0..7 (R, ΔR, Adj Ref, Reference, Excitation voltage, 530 Heater voltage, 530 Heater current, 530 Set point). The AVS47-IB command is DISx

The inputs of the VI default to -1, which means “no change”. Therefore, you can use only one or any of the inputs and leave the rest unconnected. The unconnected parameters will not be configured nor changed.

This structure eliminates the need to have five different configuration VI's for the AVS-47B. Note that we have used standard digital controls instead of the more readable ring controls. The reason is that the numeric range of the ring controls starts from 0, negative values being impossible. Any other solution seemed to be confusing.



Looking at the VI diagram, we see that all parameters are handled similarly. The digital control and (-1) are compared. If the control indicates “no change” (default), then the case structure is bypassed. In the TRUE case, a string consisting of the command, argument and a semicolon is added to the previous string. Assume, for example, that the previous string is “INP 1;” After the second case statement, the string could be “INP 1;MUX 2;”. And so on.

After the command string is complete, the string is sent to the AVS47-IB. However, before this is done, the AVS47-IB is serially polled (**poll STB status.vi**). The status must be 0 (idle), otherwise one cannot send commands to the interface. The device status can be other than 0 if, for example, the interface has been previously programmed for scanning and this procedure has not yet ended.

The “not in idle state” output is wired to a LED indicator.

4.2 Configure AVS47IB



Configure AVS47IB.vi can be used to configure one or any of the following:

Printer enabled: Enables (1) or disables (0) printing of responses to ALL queries on a parallel printer (LPT1:) which must be connected to the printer port of the AVS47-IB. This is not a useful feature in a LabView application, because the message headers must be disabled for this driver to operate, and therefore the printout would be very difficult to read and to interpret.

The AVS47-IB power-on default is “disabled” and the LabView default is “no change”(-1). Commands to the AVS47-IB are PRN0 and PRN1.

Buffer enabled: Enables (1) or disables (0) saving responses to a RAM disk or CF disk buffer. Each command line, consisting of commands that are separated by semicolons, results in one response line, where the items are separated by semicolons. Note that this driver requires the response message headers to be disabled. Therefore the response line will consist of only numbers, which appear in the same order as the commands. The buffer is probably more useful in direct GPIB applications than in LabView.

The buffer size is 64kB and the remaining space can be read using **Read and save buffer.vi**.

The AVS47-IB power-on default is “buffer disabled” and the LabView default is “no change” (-1). In 47IB: DSK0 and DSK1.

Message headers included: Enables (1) or disables (0) headers in the response messages. Default is “disabled”.

After power-on, the AVS47-IB echoes all command and query headers in the responses. For example, if the combined query is RAN ?;MUX ?, the response will be RAN 5;MUX 0. Once the message headers have been disabled, the response will shrink to 5;0.

A response without headers is more difficult to read by a human being, but it much easier to read by a computer as no parsing is required. The **Default setup.vi**, which is used inside the **Initialize.vi**, disables the headers which is a prerequisite for this driver to work. This driver does not contain any situation where the headers should be enabled.

If you want to study the operation of the AVS47-IB by using the **Discuss with AVS47IB.vi** application example, then you may want to turn the headers on by using command HDR 1.

FSM enabled: Enables (1) or disables (0) the front panel started measurements. Default is (-1) “no change”.

The AVS-47B offers a possibility to make averaging measurements by lifting the START/PRINT button momentarily. The measurement consists of a predetermined number of A/D conversions (the number depends on excitation), whose average is calculated. The results are printed, if a printer is connected to the printer port. Also, the results are stored in the RAM disk buffer *regardless* of whether buffering is enabled or disabled. The reason for this decision was to enable printing from the buffer without needing a computer for setting the parameters (keep the switch lifted for 3 seconds). The important thing to notice is that the lines in the buffer will contain response message headers. One must avoid situation where the buffer may contain lines both with headers and without them, as parsing would be difficult. Therefore, keep FSM disabled. The **Default setup.vi** in the initialization section disables the headers. The AVS47-IB command is FSM 0.

Autorange enabled: If autoranging is disabled (0), any measurement, averaging or scanning, will be conducted using the range that was selected prior to the command. The result can be valid or overrange. It is your responsibility to check the result (using any of the overload indicators that are provided by the appropriate data VI's).

If autoranging is enabled (1), the AVS47-IB will change range upwards whenever the reading exceeds 19999. If this happens when an average is being recorded, recording is re-started on the new range. Upranging can continue up to the 2M Ω range, then the reading is taken anyway.

If an underrange (reading less than 1800 counts) is encountered, the AVS47-IB changes range downwards. Averaging, if in process, is re-started on the new range. Downranging can continue to the 2 Ω range, then the reading is taken anyway.

Averages that were taken with autoranging enabled can never be invalid due to an A/D converter overload, except if the resistance exceeds 2 M Ω . The same applies also to scanning. Autoranging has been supported by the AVS47-IB also so that all resistance queries calculate the correct value regardless of the range or deviation magnifier.

AVS-47B hardware autoranging must be disabled (MANUAL ranging). In a **temperature control application, do not enable any autoranging**, as this can lead to unpredictable results.

The AVS47-IB power-on default is ARN 0 (autoranging disabled) and the LabView default is “no change” (-1).

Stabilisation delay: This is the delay between two successive autoranging operations. Its purpose is to allow the bridge some time to find a balance before making the next measurement and before deciding whether still more autoranging operations are needed. In other words, one stabilisation delay follows each autoranging operation.

A too short stabilisation delay may lead to oscillation between two ranges. A too long delay will make the system unnecessarily slow. The length of a suitable delay depends on excitation, and on whether the anticipated jumps are large (like when multiplexing between sensors).

The stabilisation delay is used only when autoranging is enabled. It has no effect if you change the range manually and then start a new measurement.

The SCAN procedure has its own stabilisation delays for each sensor. These delays can be tailored to suit the excitations used for these sensors., because the AVS-47B is slower on lower excitation ranges. See also **Configure channel parameters.vi**.

The AVS47-IB’s default value is 15 seconds. This is usually enough for all situations (although often a shorter delay is sufficient). The LabView default is “no change” (-1).

4.3 Configure addresses



Configure addresses.vi is used to set the

GPIB device address: The GPIB address can be within range 1..30. The factory default address for the AVS47-IB interface is 20.

Please note that in older AVS47-IB boxes, the GPIB address is stored in an EEPROM memory residing in the CPU. The **write count** of this EEPROM is **limited to 10000**. The write count is incremented each time you run **Configure addresses.vi** AND either the GPIB address OR the Picobus Delay Factor is not -1 (no change). Exceeding the write limit may lead to an unreparable fault (the whole computer must be replaced). Therefore, *limit the use of this VI to temporary occasions*. Under no circumstances include it in any section that may be run routinely.

Picobus address: Picobus address is the address used by the Picobus protocol to identify the Picobus instrument. Valid range is 1..15. The Picobus address is set by a DIP switch inside the AVS-47B resistance bridge. Factory default for all AVS-47B bridges is 1. You need not change this value unless you have more than one AVS-47B connected to a single AVS47-IB interface. In such a case, the two bridges must have different addresses.

After power-on, the AVS47-IB defaults to Picobus address 1 (PBA 1). Use **Configure addresses.vi** for changing this value later, if needed. Changing the Picobus address does not increase the write count, because this number is not stored in the EEPROM.

The LabView default is -1 (no change).

Picobus Delay Factor: Picobus Delay Factor (PBD) determines the end count of a simple for-next program loop, which is used to adjust the Picobus speed so that the waveform shapes in the Picobus interface line are reasonably good. The factory default value of your AVS47-IB box has been selected to suit the standard 5 m cable and the speed of the embedded PC. You should not change this value unless your cable is much longer, or if you want to insert RF filters in the Picobus line. In that case, you may need to make the delay longer. This need manifests itself by more or less erroneous

ous readings, or by failures in the remote control. Experiment with other PBD values until you find the speed limit. Select a new PBD so that it is well slower than this limit.

Changing the Picobus Delay Factor will increment the limited EEPROM write count. Refer to the discussion of the GPIB address.

4.4 Configure TS530A



Configure TS530A.vi is used to set values for any or all of the PID parameters of the controller:

Set Point: Valid range is from 1 to 42000 corresponding to 100 μ V..4.2V. However, do not use a set point lower than 1 mV.

LabView default is “no change” (-1). (NOTE: the power-on state of the TS-530A set point DAC is random).

Proportional gain: Valid range is from 0 to 15. Values from 0 to 11 correspond to gains of 5, 10, 15...60 dB. Values 12, 13 and 14 are forbidden. Value 15 short circuits the error signal to ground.

If you want to make changes in the input conditions during temperature control, or if you want to measure another sensor temporarily, then short the error signal and latch the integrator (infinite time constant).

Note that the indicated gain values are not overall gains, which depend also on the output stage and the heater. Think that they indicate only the gain steps that are available.

TS-530A power-on default is 0 and LabView default is “no change” (-1).

Integration time constant: Valid range is from 0 to 11. Values 0 and 11 both indicate the PD (proportional + differentiate) mode where the integrator stays reset. Values from 1 to 9 correspond to time constants 1, 2, 5, 10, 20, 50, 100, 200, 500 and 1000 seconds, respectively.

The integrator can be latched by setting TI=10. Latching means that the time constant is infinite. However, because the integrator uses analog technology, there will be slow drift either upwards or downwards.

TS-530A power-on default is 0 and LabView default is “no change” (-1).

Derivation time constant: Valid range is from 0 to 7 corresponding to derivator time constants of 0, 1, 2, 5, 10, 20, 50 and 100 seconds.

TS-530A power-on default is 0 and LabView default is “no change” (-1).

Power bias: Valid range is from 0 to 5 corresponding to 0, 20, 40, 60, 80 and 100% of the full power of the range. The bias power drive signal is summed to the power drive signal from the PID controller (which may be either positive or negative).

Bias power is useful in the PD control mode, where it can reduce the magnitude of the steady-state error.

TS-530A defaults to 0 at power-on, whereas this VI defaults to “no change” (-1).


Heater power range: Valid range is from 0 to 7 corresponding to OFF, 1 μ W, 10 μ W..1W.

These values apply to a 100 ohm heater. If the heater’s resistance is lower than 100 ohms, the maximum output current will limit the power (100mA on the highest range, then divide successively by 3.16). If the heater’s resistance is higher than 100 ohms, the heating power will be limited by the maximum output compliance voltage, which is about 10 Volts. Refer also to the TS-530A manual.

Power-on default of the TS-530A is 0 and LabView default is “no change” (-1).

You can use **Configure TS530A.vi** to program one or more of the PID parameters by giving values to the corresponding inputs. Those parameters that are left unwired, or whose values are set to -1 are not changed. Note, however, the way how the AVS47-IB interface box works. Whenever you change at least one of the PID parameters, all parameters will be sent to the TS-530A. This happens regardless of the fact that this VI may generate only a short command string with just one parameter command, for example.

The purpose of this behaviour is to make sure that all parameters correspond to values that *the interface* has in its memory, and that nobody can make permanent changes in the settings by using the front panel switches. In the case of the AVS-47B bridge, this has been accomplished by disabling the front panel controls, but because the old design of the TS-530, this method was not available to this driver.



4.5 Enable service requests



You can use the **Enable service requests.vi** to enable either one or both of the following events to generate an GPIB service request: Message available (MAV) or Operation Complete (OPC).

Above is the diagram of this VI. The two Boolean controls (MAV and OPC) are first combined into an array of two elements (build array). Then this Boolean array is converted into an integer number. Commands required for programming the SRE (Service Request Enable) mask register are listed upmost in the diagram as array elements (refer to the AVS47-IB manual for more information on the masks and the status reporting in general).

There are four commands corresponding to the four possible “and/or” combinations of the two inputs. The “index & append” function selects one of these elements to be sent to the AVS47-IB. Whether the command is written depends on the state of the interface. This is found by polling its status (**Poll STB status.vi**). Zero status indicates idle state and writing is possible. A nonzero value results in setting an error flag (“not in idle state”) and the VISA write operation is bypassed.

This driver does not use service requests nor interrupts. Instead, all waiting loops have been designed by using only serial polling. Then, using a delay, one has complete control over the behaviour of the system. By making the delay shorter, polling is made more frequent and the response toward the instrument is faster. By making the delay longer, polling is made more seldom, but other tasks, processes or programs that are running in the computer get more CPU time with less overhead. LabView offers a VISA function “Wait SRQ”, which could probably have been used. Unfortunately, the behaviour of this function has not yet been described in detail so we felt it safer not to use it.

4.6 Configure time and date



Configure time and date.vi takes the current time and date from the computer and initializes the real time clock of the AVS47-IB to these values. This VI uses the “Format date / time string” -function. The format strings make the outputs to suit the interface, and commands for both the time and date are combined. The combined command could look like DAY 2000,1,31;TIM 11,55,59 (31st January 2000, at 11:55:59 am).

You will need this VI if you save scanning result in the buffer. These results have time labels, which are of best use if the clock time is correct.

The clock inside the AVS47-IB is not battery-backed. Therefore it will lose its value in case of power outage. This driver version (1.2) does not provide means for detecting such a condition.

4.7 Configure scan parameters



The **Configure scan parameters.vi** is used to configure one or any of the general settings that control the SCAN procedure. These settings do not depend on the channel number being measured (the channel-specific parameters are given by the **Configure channel parameters.vi**). A few rules must be obeyed.

First channel: This is the first channel to be included in the SCAN procedure. It can be anything from 0 to 7. **Except:** If the TS-530A temperature controller is included in the system, the control sensor *must* be connected to channel 0, and this channel *must not* be included in the SCAN sequence. AVS47-IB default is 0 and LabView default is “no change” (-1).

Last channel: This is the last channel to be included in the scan sequence. It must be greater or equal to the first channel. If first channel = last channel, then this one sensor is measured using those channel parameter values that were given for this sensor. AVS47-IB default is 0 and LabView default is “no change” (-1)

Scan interval: Scan interval (SCI) is the time between two successive starts to measure the first channel. This parameter is meaningful only in the continuous scan mode. The scan interval must be longer than what is actually needed to measure all the sensors that are included in the sequence. If it is shorter, the new cycle begins immediately after the previous one. Valid range is from 0 to 10000 seconds.

Note that it is not possible to specify the exact times when the sensors are measured. This is due to at least two facts: 1) Possible autoranging operations can cause unpredictable delays and 2) In order to separate a real zero reading from an A/D converter overload, one has to repeat a couple of times a measurement that results in a zero reading. So some measurements can be randomly longer than the others.

If temperature control is enabled, the scan interval must be so long that both the scan sequence and the temperature control period have time to complete before starting the next cycle.

AVS47-IB default for SCI is 600 seconds and LabView default is “no change” (-1).

Enable temperature control: In addition to scanning sensor channels from 0 to 7, it is also possible to allocate sensor 0 for temperature control and to scan one or more of the remaining 7 channels.

If temperature control is enabled, the SCAN procedure is changed so that the temperature control is active for a period called TCP (temperature control period) between measurements of any two successive channels. Temperature control is active also for the time after the scan cycle but before the next cycle starts at the time determined by the SCI .

The combination of the temperature controller and scanning is explained in more detail in the AVS47-IB manual.

AVS47-IB default is “disabled” (ETC 0) and LabView default is “disabled” (0).

Temperature control period: This is the time used for temperature control between scanning any two successive channels. The temperature control period must be chosen experimentally. The idea is to prevent the system from drifting too far from the thermal equilibrium during the SCAN procedure, when the integrator is latched and the controller outputs a constant heating power instead of an active control.

Valid range is from 0 to 1000 seconds. If TCP=0, the controller will be active only between the ends and next starts of the scan cycles.

4.8 Configure channel parameters



Configure channel parameters.vi is used to specify the range, excitation, stabilisation delay and average count for each sensor channel separately.

You must use this VI once for each channel, it is not possible to program several channels at the same time.

The AVS47-IB requires a command string like SCP5;RAN4,EXC3;SDY10;CNT30. This string means the following: Commands that follow on **this line** are intended for sensor channel No. 5. This channel is to be measured on range 4 (2 kOhms), using excitation 3 (30µV). The stabilisation delay after having selected this channel and its parameters, and also in case of an autoranging operation, is 10 seconds. The final result shall be the average of 30 successive A/D conversions.

The diagram of the VI is very similar to those where the AVS-47B or the TS-530A are configured, except that the first entry, the channel number, is mandatory. You can leave any of the channel parameters unchanged by not connecting the inputs or by setting them to -1.

The combined command string will be written only if the interface is in idle state.

Channel: 0 to 7. If temperature control is enabled, then the valid range for channel is from 1 to 7. This entry is mandatory, there is no default.

Stabilisation delay: Any time a new sensor channel is selected, the AVS-47B bridge needs some time to settle. The same applies also to an automatic change of range. The settling time depends on the excitation - the bridge is faster when higher excitation is used. By using different stabilisation delays

for different sensors, you can optimise the speed of the scan procedure but prevent premature measurements.

Valid range is from 0 to 100 seconds.

AVS47-IB default for all channels is 15 seconds, which is sufficient even for the lowest excitation. The LabView default is “no change” (-1).

Average count: Generally, use an average count greater than unity. It is meaningful to improve the confidence of the readings by averaging, if you have time. There are many delays in the SCAN procedure anyway, so that short averages do not make the cycle much longer.

Valid range is from 1 to 1000. Each A/D conversion need 0.4 seconds.

AVS47-IB default is 10, the LabView default is “no change” (-1).

Excitation: Valid range for excitation is from 1 to 7. AVS47-IB default for all channels is 1 (3 μ V). This value was chosen in order to prevent incidental overheating of the sensors. LabView default is “no change” (-1).

Range: In manual range mode, this setting will be used regardless of the result. It is your responsibility to verify that the result is not invalid. Because this is difficult to do while scanning, use autoranging whenever possible. Autoranging guarantees that the results are always valid (except if the resistance is over 2M Ω).

In autorange mode (see **Configure AVS47IB.vi**), this parameter defines only the initial range. The initial range is tried when this sensor is measured for the first time. If needed, one or more autorangings are made until a valid reading is obtained. Then the average is recorded. The new range setting will be remembered when this sensor is measured during the next scan cycle. This means that the next scan cycles may take less time.

AVS47-IB default is 2M Ω for all channels, and the LabView default is “no change” (-1). If you use autoranging, you do not need to select any other initial range value. The highest range is a safe guess in the sense that it does not cause accidental overheating.

4.9 Configure reference voltage



The deviation reference voltage of the AVS-47B can be set remotely in two ways by using the **Configure reference voltage.vi**:

Programming mode: The default programming mode (0) is to give the desired reference voltage. The alternative mode (1) is to read the A/D converter and use this reading as the reference. The latter mode effectively means nulling the **deviation** signal.

Reference voltage: Valid range is from 0 to 2.0000 Volts. Because the reference D/A converter has only 12 bits (corresponding to 4096 counts), the allowed voltage values must be rounded to the nearest number that is divisible by 5. Therefore the step of this digital control has been set to 5.

5. ACTION/STATUS VIs

According to National Instruments' definition, the Action/Status category contains two types of VIs. Action VIs start or stop measurements. They do not change the settings of the instruments otherwise. Status VIs obtain the current status of the instrument. This driver contains only one status VI, **Poll STB status.vi**.

5.1 Go remote



Go remote.vi sends only one command to the AVS47-IB, and that is REM 1. After this command, the interface reads the current settings of the AVS-47B resistance bridge to be the initial values for the remote mode. During this transaction, the remote bit is held reset so that the bridge remains in local mode. The AVS47-IB then sets the remote bit and writes the obtained values back to the bridge. Now the bridge enters into the remote mode, and at the same time all switch settings are programmed according to their previous values and the state of the bridge remains unchanged.

As the result, **Go remote.vi** does not change the settings of the AVS-47B.

This VI is used only once, in the **Initialize.vi**.

5.2 Go local



Go local.vi sends only one command to the AVS47-IB, namely REM 0. Within this driver, **Go local.vi** is used only once, in the **Close.vi**.

Usually you want to leave the system in the local control state so that the front panel switches are active. If you want to prevent anyone from changing the settings, you can also close your LabView application so that the AVS-47B is left in its prevailing mode, local or remote. Refer to **Close.vi**.

5.3 Poll STB status byte



This is the only status VI in this driver. **Poll STB status byte.vi** can be used regardless of the state of the AVS47-IB. This is because serial polling is handled by the 7210 GPIB controller quite independently of the embedded PC.

The application program in the AVS47-IB updates a byte which we call “serial poll response **spr**” whenever the program moves from one phase to another. The **spr** byte is maintained in a register of the 7210, which can therefore respond immediately to a serial poll at any time. The information of the **spr** is also available in the STB master status byte (*STB? query), but only when the system is idling. In this manual, we use only the term STB, even if it is obtained by reading the serial poll response. The difference between these two bytes is the service request bit, which is nulled by reading in **spr** but not in the STB (refer to AVS47-IB manual).

The STB (or **spr**) of the AVS47-IB contains the following information. When reading the rather complicated description, please note that we have done all the necessary operations for you. Most probably you will not need to use this VI yourself at all.

bit 2⁶ SRQ. The system is requesting service. This bit can be set only if one or more events have been first enabled, and any of these enable events occur. Refer to **Enable service requests.vi**. The service request is a physical signal line in the GPIB bus. It is directly wired together with this SRQ bit.

Usually, GPIB controller cards offer a possibility to generate a hardware interrupt from the SRQ signal line. This is one way to detect an SRQ.

It is also possible to repeat serial polling and inspect bit 2⁶ until it indicates an SRQ. This is the second way to use SRQ.

This driver does not use SRQ, because the few events that are important can be directly seen from the STB status byte by serial polling.

Serial polling the STB resets this bit (it is required by the IEEE-488.2 standard), which means

that SRQ can be detected only once per event. No event can generate a new SRQ unless the Event Status Register is first reset with *CLS.


bit 2⁵ OPC, operation complete. An OPC event can set this bit only if appropriate mask register bit has been enabled. This is done in the **Initialize.vi**.

In order for the OPC event to appear, one must first clear the Event Status Register and also place a *OPC command as the last item in the command line.

bit 2⁴ MAV, message available. This bit in the STB indicates that the output buffer of the AVS47-IB is not empty. As the interface will never produce any output by itself, MAV=1 can only mean that the response to a query is available.

bits 2⁰..2³ This group of four bits (16 values) describes the state of the interface. You may not need to use the **Poll STB Status.vi** yourself, but nevertheless we describe the status signals, as this can help in understanding the behaviour of the instrument.

- 0: Idle state.** The 47IB accepts all commands and queries. Idle state means that the unit is not measuring, not scanning, not controlling temperature neither doing anything else. It is just idling.
- 1: Averaging.** The unit is making A/D conversions at maximum speed for calculating their average in the end. Only two commands are allowed: STP (stop) and *RST (reset).
- 2: Scanning.** During some phases of the scan procedure, the unit may also show state No. 1. Only STP and *RST commands are allowed.
- 3: Scan interval delay.** The unit is waiting for a new cycle to start in the continuous scan mode. The temperature controller is active if temperature control has been enabled. Only STP and *RST are allowed.
- 4: Temporary temperature control during scanning.** This means the short temperature control periods (TCP) between measurements of successive channels. Only STP and *RST are allowed.
- 5: Waiting for the START/PRINT button in the front panel started scan mode.** You should never see this state in a LabView application! *Front panel started measurements are incompatible with this driver.* Only STP and *RST are allowed.
- 6: Digital filter in progress.** A/D conversions are being taken at maximum speed and the readings are used to fill a circular buffer. It is allowed to make a DFR? (digital filter reading) query, and to stop (STP) or reset (*RST) the interface.



- 7: **Printing the buffer on printer.** The interface goes into this state after a PBF command (print buffer file), and it exits back to state 0 once printing is complete. Only STP and *RST are allowed.
- 8: **Reading the buffer in RBF mode.** The interface goes into this state after an RBF command, and remains there as long as there are lines to be read or until reset. It is allowed to make the NXT? query for receiving the next data line, or to give the STP or *RST commands.
- 9: **Self-Calibration in progress.** Only STP and *RST are allowed.
- 10: **Self-test is in progress.** This time-taking procedure cannot be terminated. No commands are allowed. Use with care!

- 3) *OPC When the command interpreter of the AVS47-IB encounters this command, it sets bit 2⁰ of the Event Status Register (ESR). The **Initialise.vi** has already enabled this OPC bit to turn on bit 2⁵ of the status byte STB register (see also **poll STB status.vi**).

Once this command string has been sent to the interface, the VI continues with a short delay loop (10ms). This delay is very important: Suppose that the OPC status of the interface was true before the ADC command. If LabView was permitted to continue very quickly by starting to wait for the OPC to come true, it would be satisfied immediately. Some time must be allowed for the interface to reset its status before proceeding (PCs are usually much faster than the embedded CPU of the AVS47-IB box).

Note that you must verify that this operation has completed before trying to read its result. There is the **Wait for OPC/MAV/TMO.vi** written for this purpose.


5.4 Start single ADC



The **Start single ADC.vi** is used to make one measurement without any averaging. Note that if the system is in autoranging mode and the initial range is not correct, one or more conversions followed by stabilisation delays may be needed until a valid reading is obtained. Therefore one cannot know how much time this will VI require.

Like with almost all commands in this driver, the idle state is first verified. The command string consists of

- 1) *CLS Clears (resets) the status register so that it can reflect new events
- 2) ADC Tells the AVS47-IB box to wait for the completion of the **next** analog-to-digital conversion and to give its result. This command should need a maximum of 0.4 seconds.



5.5 Wait for MAV/OPC/TMO



The **Wait for MAV/OPC/TMO.vi** is the general waiting loop in this driver. It can be programmed to wait one or more of 1) **Message available**, 2) **Operation complete** or 3) **Timeout**.

The MAV and OPC are Boolean controls. If both are enabled the VI waits for the first of them to occur. If both are disabled or if they fail to occur, the VI waits only for the timeout.

The timeout value (in seconds) is set by a digital control. This waiting loop is not used for generating millisecond delays. Therefore the 1 second step is adequate. The timeout is generated by recording the computer time before entering the while-loop, by adding the timeout number to this reading, and by comparing the sum against future computer time while the loop is running. If you set the timeout control to zero, the logic prevents the VI from generating the Boolean “Timeout” error signal.

The loop delay is an important parameter. During this delay, the while-loop gives its time co-operatively to other simultaneous tasks. The longer is the delay, the better other tasks can run during this loop. On the other hand, the MAV and OPC status will be checked more seldom. The result may be that your application becomes too slow. This is a tradeoff that you have to make.

The loop delay can be turned off by setting the timeout value to 0.

5.6 Start averaging



The **Start averaging.vi** begins a series of A/D conversions whose average is calculated and can be queried using a separate VI.

This VI needs one input parameter, which is the average count. Otherwise, the measurement is made with those settings that were valid before this command. The average count can range from 1 to 1000. Each conversion needs 0.4 seconds. Taking an average of 100 conversion needs 40 seconds, so that this is not a very fast procedure.

The combined command string is formed from *CLS (for resetting the status) followed by AVEx where x is the

average count, and ending by *OPC. The last command gives us the possibility to wait until all A/D conversions have been made.

There is a short delay after this command, before the VI exits. This delay prevents any attempts to read the STB status before the AVS47-IB has had enough time to update it.

5.7 Start digital filter



The **Start digital filter.vi** sets the AVS47-IB into a special mode, where the selected channel is read as fast as possible (0.4 sec/sample) using the range and excitation settings that were valid before starting this mode. The readings are maintained in a circular buffer and read using a separate data VI. **Start digital filter.vi** has one input, the filter length (DFL). Valid range for the filter length is from 1 to 1000. You can probably find values from 10 to 100 to be most useful.

If the 47IB box is in digital filter mode, it is not idling. Polling the STB status will return 6.

Most commands/queries cannot be used in the digital filter mode. VI's that expect the AVS47-IB to idle are just bypassed and their “not in idle state” error flag is set. The only two VIs that will work are


“**Read digital filter.vi**” for obtaining one reading from the circular buffer. Refer to the description of this VI.

“**Stop.vi**” for exiting the digital filter mode. Refer to the description of this VI below.

In addition, you can always use the “**Poll STB status.vi**” to find out the present state of the interface.

The digital filter mode is a fast way to get high-quality readings quickly, if you can keep it running. Use a long filter if the excitation is low (random noise is reduced by sqrt(DFL)). The filter function is a simple running average. When the filter is started, after an autoranging operation or after a large sudden change in the reading, the filter is reset. Then the reading is the sum of the cumulated results divided by their number, until the buffer is full.

Because of its nature, the digital filter mode is not suitable for multiplexing. In such case, use normal averaging, or even better, use the SCAN procedure.



5.8 Stop averaging etc.



The **Stop.vi** is used to stop the following operations:


- digital filter mode
- averaging
- scanning, especially the continuous scan mode.
- self-calibration procedure

The only situations where stopping should be routinely needed are exiting from the digital filter and continuous scan modes, because these modes never stop by themselves.

Stop.vi can also be used to abort printing etc. (see also the description of **Poll STB status.vi**).

The VI starts with polling the STB. The status is checked to be within the range from 0 to 10 (see **poll STB status.vi**). If the system is in a state that can be stopped, the STP command is sent to the interface. In the negative case, the VISA operation is just bypassed.

The following delay is very important. It gives time for the AVS47-IB to stop the operation and to update its status byte, which is used for response to a serial poll. If the LabView application continues by reading the status before it has been updated, an error situation will result.



5.9 Start single scan



Start single scan.vi initiates one scan cycle. After all phases of the cycle have been completed, the interface returns to the state that was valid before starting to scan. It means the original channel, range and excitation. A scan cycle can be started only when the AVS47-IB is in idle state. So this is verified at first.

This VI has one Boolean input parameter, Buffer storage, which is disabled by default. When buffer storage is *disabled (false)*, you must read the results after the scan cycle has ended. The **Read scan results.vi** has been designed for this purpose.

If buffer storage is *enabled (true)*, you can read the scan results also later from the buffer. Because results of new scan cycles are appended to the buffer, you can read results of several cycles at one time. The results contain time stamps, which are useful, if the time of the non-battery-backed clock of the AVS47-IB box has been set.

The combined command string consists of

- disabling or enabling buffer storage (DSK 0 | 1)
- clearing the interface status (*CLS)
- starting a single scan (SCN 0)
- setting the OPC indicator bit in the STB when the cycle is complete.

A short delay after the VISA operation guarantees that the status cannot be polled before the 47IB has had time to reset it first.

5.10 Start continuous scan




The **Start continuous scan.vi** differs from the **Start single scan.vi** only in one place: the SCN 0 command is replaced by SCN 2.

When using continuous scan, **it is almost necessary to enable buffer storage**. This is because you *cannot* ask scan results while the system is busy. It is probably better to use single scan cycles rather than to stop the continuous

mode. If the continuous mode is stopped during a cycle, the scan results may belong to different cycles.

Do not use the **Read scan results.vi** after the continuous mode. Instead, use the **Read and save buffer.vi**.



6. DATA VI'S

According to their definition, data VIs transfer data to and from the instrument. Data VIs for the AVS47-IB all belong to the latter category, there is no data that could be sent to it (except data which is used for configuring it).

The general rule when trying to read anything from the AVS47-IB (or from any other GPIB instrument, for that matter) is that one must always verify there is something in the output buffer. Otherwise, the GPIB bus will hang up until it decides this is a timeout.

Many AVS47-IB operations take a very long time in the computer scale. One A/D conversion needs 0.4 seconds, taking an average of 100 readings needs 40 seconds, and a longer scan cycle with temperature control enabled may take some tens of minutes. It is necessary to use waiting loops which continuously poll the interface status, but which also give all their spare time to other tasks. If not done so, the computer would be useless during all this time.


The waiting loop that has been used in most data VIs is the **Wait for OPC/MAV/TMO.vi**.

6.1 Read single ADC



The **Read single ADC.vi** is intended to accompany **Start single ADC.vi**. However, you cannot connect these two VIs directly after each other, but you must first verify that the previous ADC operation has completed. Use the **Wait for MAV/OPC/TMO.vi** for that purpose. The ADC conversion result can be read only if the system is in idle state.

The AVS47-IB will receive the following command string: `*CLS;ADC ?;RES ?` At first, the box resets its status. The next query obtains the result of the last A/D conversion as a signed integer, without any knowledge on range or display mode. Note that a delay between `*CLS` and `ADC?` is not needed, because we do not try to poll the



status. The RES? query obtains the A/D conversion result as a *resistance*. The resistance is calculated by the AVS47-IB on the basis of the range and the position of the ΔR magnifier switch. Resistance is a floating point number.

NOTE 1: If the display selector is not within 0 and 4, the resistance readings are meaningless. Only the ADC? query shall be used for reading the excitation voltage, heater current etc.

NOTE 2: Because of the limitations of the hard-wired AVS-47B circuitry, the decimal point on the LED display of the bridge is connected to the $\Delta R \times 10$ switch. If the switch is in the magnifier mode, but the display selector is not in the ΔR position, the decimal point will be in a wrong place on the display. This driver calculates the result correctly, however. The same applies also to the AVS47-IB interface, which works correctly regardless of the switch settings.

6.2 Read one reading




The **Read one reading.vi** is a self-contained and relatively “easy-to-use” VI for getting one reading from the channel that has been previously configured using other VIs. This is a typical data-VI in that it does not have any inputs at all. On the other hand, the VI contains commands that start an A/D conversion, and these commands would belong to an action-VI. In fact, **Read one reading.vi** is a combination of both the **Start single ADC.vi** and **Read single ADC.vi** source codes.

If the interface is in idle state, its status is first cleared (*CLS). Then an A/D conversion is started (actually, the AVS-47B has a free-running A/D converter, which cannot be started. Instead, the ADC command instructs the interface to wait for the next conversion to complete and to use its result). The *OPC command requests the AVS47-IB to set the OPC indicator bit in its status register when the conversion has completed.

A short delay prevents the VI from proceeding before the interface has had time to reset the status byte.

By setting the OPC control *true* and the MAV control *false*, the **Wait for OPC/MAV/TMO.vi** is instructed to wait until the end of the conversion. If the conversion does not complete before the default timeout of 2 seconds, an error flag is set, and the rest of the VI is bypassed (such a situation could result from a missing or faulty Picobus cable, from a loose optical fibre in the **Picolink**, or if the AVS-47B is simply off).



The status is cleared again, and the interface is queried for both the integer ADC result and the floating point resistance value.

No delay is required here before entering the waiting loop. This is because the status was cleared once before starting the A/D conversion, and the conversion does not produce any output by itself. So the “message available” -bit cannot be high when the waiting loop starts. If the response does not appear within 2 seconds, the VI exits with the timeout error flag true.

The VISA read -function obtains the integer ADC result and the floating point resistance reading. Note that some AVS-47B display items do not deal with resistance (items 4-7). Then use only of the ADC result and scale it.

The ADC result is compared against 20001. This value, used to indicate overload, is beyond the range of the A/D converter. There is also a separate OVL ? query, but using 20001 is faster as no additional commands are needed.

Note that the resistance has been calculated by the 47IB box: it takes into account the range and the position of the 10*ΔR switch. The reading is correct regardless of the position of the magnifier switch.

6.3 Read average



The **Read average.vi** returns, in addition to the average value itself, also the minimum, maximum and standard deviation of all recorded A/D conversions.

The **Read average.vi** has not been intended to be used with display items other than R or ΔR. The average


is a floating point number, corresponding to the resistance in all other respects, but it gives better resolution because random noise is reduced by averaging.

This VI works as follows: The **Wait for OPC/MAV/TMO.vi** is set to wait for a completed operation, with a timeout of 500 (five hundred) seconds and a loop delay of 500 milliseconds. The timeout value must be really long, so that this VI can be started immediately after having initiated averaging. An average of 1000 readings (this is the maximum) will take 400 seconds.

The loop delay is also long, 500 milliseconds, so that as much time with as little overhead as possible can be given to other tasks.

If not timeout, the above mentioned readings are queried, followed by the overload query OVL?. Note that here we must use the overload query, because it tells if any of the averaged A/D conversions resulted in overload.

The next WaitFor loop delays operation until the response is ready to be read. The VISA read function receives a string consisting of four floating point numbers followed by the overload indicator which can be only 0 or 1. All items in the response are separated by semicolons.



6.4 Read digital filter




The digital filter reading can be obtained using the **Read digital filter.vi** only if the AVS47-IB is in state 6 (digital filter in progress). Therefore the state is first verified by polling the STB status byte.

The status is cleared so that one can wait for MAV and the filter reading is queried by DFR?.

The response is read once it is available in the output buffer of the interface box.

The filter reading is a floating point number, corresponding to the “resistance” and “average” values. Digital filter is not suitable for measuring display items such as excitation voltage or TS-530A set point, where range setting or the position of the ΔR switch must not affect the reading.

Do not use digital filter together with multiplexing. Use, for example, the SCAN procedure with first channel = last channel.



6.5 Read scanresults



The **Read scan results.vi** should be used only for reading the results of a single scan cycle. It will not be useful for the continuous scan mode. One must stop the continuous mode for asking the results, but after stopping, the results in the memory may belong to different scan cycles. That does not happen with single scan cycles.

This VI has been designed so that it can follow directly after the **Start single scan.vi**. Compare this VI with the **Read average.vi**, and you see many similarities but also some differences.

The first thing to do is to wait until the scan cycle completes. This is made by programming the **Wait for OPC/MAV/TMO.vi** to wait for OPC with a long timeout (500s) and a generous loop delay (500ms). Refer to the discussion of **Read average.vi**.

This VI needs one input parameter, which is the channel number. You must ask the results for each sensor separately by varying the channel number. The command that tells, which channel to read, is SCR x , where x is the channel number. The total command string starts with clearing the status, then entering into the reading mode, followed by commands for reading everything that is available: *CLS;SCR x ;AVE?;MIN?;MAX?;STD?

The response is read **once it becomes available** in the output buffer of the interface. If you use this VI after having started a single scan cycle, the VI waits until the cycle has ended and the results are asked immediately after that. You have to repeat this VI for each channel whose data you want to read.

When doing so, you must necessarily disable the “wait for completion” for all other than the first VI. Otherwise, as there is no longer an OPC bit to watch, the VI would wait until timeout for each subsequent read.

The response string consist of four floating point numbers, separated by semicolons.

Note that overload has not been queried nor detected from the output. We want to strongly encourage you to use autoranging instead of manual ranging together with scanning. Then all results will be valid.

6.6 Read and save the buffer



The purpose of the **Read and save buffer.vi** is to enable reading and saving the buffer as a disk file after the AVS47-IB has been operated in the continuous scan mode. It cannot be used for other purposes because this VI is dependent on the unique data format used for buffering the scan results. Reading is possible if the system is in idle state.

This VI is more complicated than those described so far in this manual. Please refer to the diagram of the VI by opening it in LabView.

After having verified that any possibly pending operation has ended and that the system is in idle state, size of the buffer and the remaining free space are queried by BUF ?. Note the DSK0 command that precedes BUF?. We do not want the response of the BUF? to go into the buffer, mixing its structure. The response consists of two numbers, the first of which is the size in bytes while the second number tells the free space. The numbers are separated by a semicolon.

The VI requires two inputs, Mode and Filename. If Mode is 0, only the buffer size and the free space are queried and then the VI exits. If Mode is 1 AND the buffer size is greater than zero, the buffer is read and saved in Filename. If Mode is 2, the buffer file is also deleted before exiting the VI.

The interface enters in the “read buffer” mode when it receives the RBF (read buffer file) command, although nothing happens yet, except that the state of the AVS47-IB is now 8. The oldest line in the buffer is placed in the output queue by the query NXT ?

The most important part of the VI is a while loop. This loop starts with a short delay. The STB status byte is polled, and if the state is 8, which means there is still something to read, one line is read from the interface. This line is then parsed and converted into a form, which might be called “spreadsheet form”. It means simply that the numbers are separated by blank spaces. Such a file can be easily read into a spreadsheet program like Excel. Once the line is in correct form, it is appended to the output file. Then the NXT ? query is repeated and the loop starts from the beginning with the short delay.

The AVS47-IB stores lines in the buffer as follows:

```
1993,12,6;4,37,24;1.0006E+02;1;0;3;7;0;0
```

Which *simply* means the following: this line was saved into the buffer on 6th December 1993 at 4:37:24 am (24-hour clock). The resistance value was 100.06Ω , the input switch was in MEAS position, channel 0 was selected, range was 3 (200 Ω), excitation was 7 (3mV), the display selector was in “R” position and no overload was detected.

These are converted to the “spreadsheet” form in a for-next loop. This loop operates as follows:

A new complete line from **VISA Read** enters into the shift register of the for-next loop. The “Match Pattern” function searches for the first occurrence of either a comma (,) or semicolon (;). Once found, it places the part before the search string into a string concatenating block.

The part after the search string is placed into the shift register where it replaces the input to the Match Pattern function.

The string concatenate block consists of three elements. The first element is an empty string in the beginning. The middle element is the output from the Match Pattern. The third element is a space character (\s) except when the loop ends.

The first of the two shift registers handles the rest of the line, which shrinks after each turn. The second shift register gathers the converted string, where the commas and semicolons have been replaced by blanks, and the contents of this shift register grows after each turn. After 13 turns (the turns count goes from 0 to 12), the second shift register outputs its contents to the Write File function, which appends the new line to Filename (must include the complete path). Now the next data line can be read from the interface.

There is no default for the Path and File names, so you have to specify them according to your computer and application.

In normal operation you should always delete the buffer file after having read it. The possibility to read and save without deleting exists for the purpose of debugging and developing the application. It would be very inconvenient to generate new buffer files frequently, because the only way to do so is to start new scan cycles.

The command used for deleting the buffer file is KIL. A deleted file cannot be recovered.


6.7 Read AVS47 parameters



The **Read AVS47 parameters.vi** is used to determine the positions of all the AVS-47B front panel switches. These positions can then be used as initial values for later remote control.

The AVS47-IB requires one Picobus transaction in order to ask one parameter. This VI queries routinely all 8 parameters and therefore it is quite slow. It should not be necessary to use it repeatedly in a remote control application.

National Instruments seems to think that a LabView driver does not need VIs for reading any settings, because these should be known by the computer in a remote control application. However, we have designed this driver so that an application can be started without disturbing a running system. This is possible only, if the current state of the system can be read.



The **Read AVS47 parameters.vi** gives the following outputs, which all are integer numbers:

- **Remote: 0..1:** Local/Remote control mode
- **Input: 0..3:** Zero/Meas/Cal
- **Channel: 0..7**
- **Range: 0..7:** no range, 2Ω...2MΩ
- **Excitation: 0..7:** no exc, 3μV...3mV
- **Display: 0..7:** R, ΔR.....530 set pt
- **Ref voltage source: 0..1:** from memory, from helipot
- **Deviation magnifier: 0..1:** ΔRx1, ΔRx10

All these items are separate outputs, so you can select which of them to use. **Getting started.vi** shows how the instrument state can be read and turned into initial conditions for later remote control.

6.8 Read from TS530A



The basic design of the TS-530 Temperature Controller dates back to a time where the computer control was limited to few aspects only. Therefore the TS-530A does not contain an A/D converter of its own. Voltages that are proportional to the heater output current and voltage, plus the analog set point voltage generated by the set point DAC, are all fed to the AVS-47B resistance bridge. The A/D converter of the bridge can then be used to measure these quantities.

The AVS-47B display selector positions 5-7 exist for this purpose: The set point voltage can be directly measured and the value displayed by the AVS-47B is correct. The heater current and heater voltage displays are not correct as such - they need some calculations which take into account the power range setting.

While the AVS-47B display cannot be directly used for measuring the heater current or voltage, the AVS47-IB interface makes all the necessary calculations. In addition, it offers also a query for the heating power.


The **Read from TS530A.vi** uses the macro commands offered by the interface unit. Because these macros need a couple of A/D conversions with settling times between them, their operation is very slow, typically a couple of seconds per query. Therefore this VI was not designed to provide answer to all four queries automatically. You must first specify the input parameter “Quantity to Measure”, which can be

- Set point (0: default)
- Heater voltage (1)
- Heater current (2)
- Heater power (3)

The floating point output is either volts, amperes or watts depending on the input parameter value.

The VI operates as follows: After having verified the idle state, measurement of one of the four items is commanded by using a string that begins with *CLS for clearing the status and ending with *OPC for generating the OPC bit once the measurement is complete. The “Index & Append” function of the LV is used to select one of the command strings.

A delay follows the VISA write, so that the interface has time to reset its OPC status before LabView starts to poll it. The **Wait for OPC/MAV/TMO.vi** loop is set for OPC, timeout of 10 seconds and loop delay of 200 milliseconds, which should allow ample of time for multi-tasking.



Once the OPC status has been detected, the corresponding item can be queried. Note that these macros are always used in pairs, like SPV;SPV? - a measurement command and the corresponding query. The status is cleared again, and this time the waiting loop, in its default condition, waits for the response to become available (MAV).

Application example 1.vi shows one possible way to use this VI.

6.9 Read TS530A parameters




It is not possible to read the PID parameters from the TS-530A in the same way as they can be read from the AVS-47B resistance bridge. This is due to the old design of the controller. However, it is possible to read the PID settings from the AVS47-IB interface. Although not exactly as useful, also the parameters stored in the interface can serve as good initial values for later control. Generally, these settings are the same as those in the TS-530A,

except in a case where none of the parameters has been remotely programmed after they have been changed manually.

The **Read TS530A parameters.vi** is similar to the read VIs for the AVS-47B and for the AVS47-IB parameters. An array of strings is first formed from all the necessary queries, in this case SPT?, PRO?, ITC?, DTC?, BIA? and POW?

The for-loop that follows selects one of the queries in turn and sends it to the interface. Once the response is available, which is verified by the WaitFor OPC etc. VI, it is read. Note that the for-loop in LabView works so that if a turn of the loop produces some output, these outputs are collected together into an array. When the loop finally exits, it is this array that is passed as output. Therefore, the output needs no parsing. We have used several Index & Append blocks in parallel to extract each element.



Read channel parameters.vi

6.10 Read AVS47IB parameters



The **Read 47IB parameters.vi** works exactly in the same way as the previous VI, except that the queries are now PRN?, DSK?, IBA?, PBA?, PBD?, HDR? and FSM? corresponding to

- printer enabled
- buffer storage enabled
- GPIB device address
- Picobus address
- Picobus delay factor
- message headers included
- front panel started measurements enabled

Note that mere querying the IBA or PBD does not increment the EEPROM write count.

6.11 Read channel parameters



The SCAN procedure uses separate sets of parameters for each sensor channel. These parameters, which are called channel parameters, are the stabilisation delay, average count, excitation and initial range. Channel parameters can be read for one channel per one instance of the **Read channel parameters.vi**. Reading is possible when the system is in idle state.

The command string that is sent to the interface is

*CLS;SCPx;SDY?;CNT?;EXC?;RAN?

*CLS resets the status byte. SCPx selects the desired channel number x, where x is the input to this VI. Then the four parameters are queried.

The response is read once it becomes available. It consists of four floating point numbers, which are extracted using the “from decimal” LabView function.

You may need this VI only for creating correct initial values for the LabView application by reading them from the interface. Whereas the standard range and excitation values can be changed from the AVS-47B front panel in the local control mode, it is not possible to change the channel parameters manually.

6.12 Read scan parameters



The **Read scan parameters.vi** is exactly similar to **Read TS530A parameters.vi**, so please refer to its description. The items that are queried are 1) the first scan channel, 2) the last scan channel, 3) scan interval, 4) temperature control enabled, 5) temperature control period.

You may need this VI only for creating correct initial values for later LabView control.

7. UTILITY VI'S

According to NI's definition, Utility VIs perform a variety of operations that are auxiliary to those VIs of the instrument driver that are used mostly. An example of a utility VI is the **Reset.vi**.

This driver version does not contain some VIs that should belong to an instrument driver if it were to meet the National Instruments' "standards". Please see "Omitted VIs" below.

7.1 Reset



The **Reset.vi** sends only one command to the AVS47-IB, namely *RST. This command can be sent to the interface regardless of its state with the exception of the self-test procedure, which cannot be terminated.

Using the **Reset.vi** will bring the AVS-47B and the TS-530A into their reset states. Most data in the AVS47-IB interface will remain intact, including the self-calibration data and scan parameters. The buffer file is not deleted.

Although **Reset.vi** exists as a mandatory option for the **Initialize.vi**, it has not been used in this driver because of its catastrophic effect on a running system.

7.2 Default setup



The **Default setup.vi** sets the AVS47-IB into a state where it is compatible with this driver. The command string is:

```
FSM0;HDR0;*ESE1
```

which means that front panel started measurements are disabled and so are also response message headers. Bit 2^0 (the Operation Complete Bit OPC) of the Event Status Register ESR is OR'ed (enabled) into bit 2^5 (the Event Status Bit ESB) of the STB status byte. This is accomplished by programming the Event Status Enable mask register ESE using the *ESE1 command. OPC is then the only event that is enabled into the ESB bit. Therefore this bit tells whether an operation is complete, whereas bit 2^4 (the message available bit MAV) is inherently enabled.

These arrangements belong to the IEEE-488.2 standard. If you want to know more about how the AVS47-IB reports its status, please refer to the AVS47-IB manual and to the IEEE-488.2 standard.

The **Default setup.vi** is used only in the **Initialize.vi**.

7.3 Self test



The **Self test.vi** initiates a thorough investigation of the system. Please refer to the AVS47-IB manual for the detailed description of the self test.

The self test procedure includes the Picobus connection, the remote controllability of the AVS-47B and the remote controllability of the TS-530A. Further, it checks the offset and scale factors of the AVS-47B and also those of the TS-530A set point DAC.

In order to use the self test, you need only to connect the **Initialize.vi**, the **Self test.vi** and the **Close.vi**. There is no verbose report - only a return number, which is called self test error, is output - but explanations can be found in the interface manual.

The self test needs some minutes to complete. If you do not have a temperature controller in the system, the possible results will always be greater than 64.

*Note that one cannot use the **Discuss with AVS47IB.vi** for starting the self test query. This is because **Discuss** has a 1 second timeout whereas the self test takes some minutes.*

7.4 Self calibrate



The **Self calibrate.vi** makes a digital calibration of the AVS-47B. The zero offsets and scale factors are measured in the ZERO and CALIBRATE input modes for all seven excitation ranges. These calibration factors are used to correct the following readings

- resistance, ΔR and $\Delta R \times 10$
- average (also in the scan mode)
- maximum (also in the scan mode)
- minimum (also in the scan mode)
- standard deviation (also in the scan mode)
- digital filter reading

The A/D conversion result is not modified, neither are the measurements from the TS-530A. **Note that digital calibration does not change the display of the AVS-47B.**

The **Self calibrate.vi** has three operating modes

- 0 (default)** Query the self calibration factors. This produces a string output, consisting of seven offsets and seven scale factors separated by commas.
- 1** Reset self calibration factors. All offsets are set



to zero and all scale factors are set to 1. After this operation (or if self calibration has not yet been conducted) the output displayed by the AVS-47B and the digital results from the interface are the same.

2 Self calibrate.

The self calibration factors are stored in the AVS47-IB as long as it is powered. If you need to check whether the calibration still exists, you can check the offset or scale factor corresponding to the $3\mu\text{V}$ excitation. Because of noise, these readings are sure to deviate from exact 0 or 1.

Use self calibration if the AVS-47B is subject to large changes in the ambient temperature. Digital calibration is faster and easier than the analog calibration which requires opening the instrument and adjusting a number of trimmers.

You can also use digital calibration to calibrate the deviation mode ΔR . This is because the calibration procedure handles the deviation simply as an offset.

Calibration factors are stored in a volatile memory. After power-on a new self-calibration is needed.

8. OMITTED VIs

Some VIs have been omitted from this driver, although they should belong to an instrument driver, according to National Instruments “standards”.

Revision Query.vi

The revision number and other necessary information can be found in “Show VI info” for the **VI-Tree.vi**.

Error Query.vi

Many virtual instruments in this driver contain error flags for “not in idle state”, “timeout”, and sometimes also for “overload”. If the interface was not in idle state, the rest of the VI is usually bypassed. The VI front panels contain error indicator LEDs, which can be inspected by opening the VI.

Once the application has been designed and debugged, error situations should be rare.

Error Message.vi

Was not implemented because no error query exists.


9. APPLICATION EXAMPLES

The previous discussion of the virtual instruments concentrated on how these VI's operate and what they actually do. Sometimes the discussion was quite complicated, because the operations **are** complicated. Very little was said about how to use these VIs simply and effectively, and this is just the purpose of LabView.

We have designed a couple of application examples, or "instrument front panels" if you want to say it that

way. These examples should give some idea on how we thought our driver could be used. We have also played with these "non existing instruments" and they seem to work in a way that we think is correct.

Even though you may want to write your application yourself by using the more basic VIs, please take a little time, read through the descriptions and try some of the examples before starting your work. And if your own VI stops, check that you have a required delay and that you wait for MAV or OPC.



9.1. Discuss with AVS47IB.vi



This is not actually an application example at all. The **Discuss with AVS47IB.vi** is a simple VI that you can use for sending commands and queries, single or combined, to the AVS47-IB interface and to read the response if such is generated.

You can use this VI when you start to use our GPIB interface box with LabView for the first time. Some operations, like changing the GPIB address or asking the firmware version, are easier and faster using **Discuss** than if you write a VI, which may be never needed again. **Discuss** provides also an easy way to study the behaviour of the AVS47-IB + AVS-47B + TS-530A system and how they react to GPIB commands.

The VISA session is opened at first. The resource name is GPIB::xx, where xx is the GPIB device address that you give from the front panel. The factory default address of the AVS47-IB is 20.

You can change the GPIB address once you have first established the connection using the old address. Then issue the command "IBA newaddress". The response to the "IBA?" query has changed



to "IBA newaddress", but the interface will still use the "oldaddress" as its GPIB device address. Boot the interface by plugging the power off and on again.

After booting, the new address becomes valid everywhere, and you must change also the GPIB address on the Discuss front panel.

There is a "property node" after the VISA open function. It seemed to us that some error conditions were avoided by adding the "GPIB readdressing=default=true".

The command line, which may consist of several commands and/or queries separated by semicolons, is written to the interface. A delay loop prevents polling of the status until one second has passed. If a message is available, it is read and placed to the response screen on the front panel. If there is no message, an empty string is placed to the response screen.

If everything works, this VI is easy to use: Open it, write the message, run the VI and read the response. Do not try Discuss with the *TST? query.



9.2. Getting started.vi



The **Getting started.vi** shows one way to use the virtual instruments in this driver to build an easy-to-use resistance bridge without any bells and whistles.

The resource name for the **Initialize.vi** has been generated from the header string and the GPIB address that is obtained from the front panel. The integer number is converted into a string using “To decimal” function, whereafter the header and the address are added using the string concatenate function.

The **Read AVS47 parameters.vi** queries the settings of all AVS-47B front panel controls. The outputs are placed in local variables having names like “inp” or “mux”. This is the way how one can read the existing setup of the system into LabView.

After the settings have been saved, the program enters into a while loop, which is run until the “STOP” button is pressed. At first, the system is set into remote mode. You can see that the yellow “remote” led on the AVS-47B front panel lights. Actually, the **Go remote.vi** could have been before the while loop as well.


The AVS-47B is now configured. All the real switches on the AVS-47B front panel have their counterparts as digital ring controls. These ring controls have a verbose name for each digital number, which makes their use easy. When the while-loop is run for the first time, values for these ring controls are obtained by copying them from the local variables having corresponding names.

In fact, LabView has supported creating such local variables. Using the wiring tool, point the digital control “inp”, for example. Right click, and you can find a menu item Create/local variable.

Once the controls have so got their initial values, you can start to change them. The local variables are used only once, so your changes remain valid as long as the while loop runs.

The **Wait for OPC/MAV/TMO.vi** is configured to produce a programmable delay. You can experiment with this delay. Set it to zero, the excitation to 10 μ V and switch between input=0 and input=100R calibrate. If the delay is zero, you get many readings before the bridge has settled. If the delay is long, the bridge has time to stabilize but the system is slow. The purpose of this delay is just to show that you need a delay after the settings have been changed.

The **Read one reading.vi** outputs both the resistance and integer ADC values.



9.3. Application example 1.vi



This is a more complicated VI, which suggests how the lower level VIs can be used programmatically to create a new non-existing instrument. Some useful hints may be found from this VI; also one peculiarity of LabView 5.01 is discussed.

The **Application example 1.vi** combines both the AVS-47B and the TS-530A into a single precision temperature controller for dilution cryostats. It can also be run without the controller, but not without the bridge. This VI offers only manual ranging. Further, you must select the control polarity of the TS-530A manually - it cannot be controlled remotely.

The AVS-47B switch panel on the left side has controls for the input mode, channel, range, excitation and for the GPIB device address. Control for the display item is found under the LED display.

The TS-530A switch panel offers controls for the proportional gain, integrator and derivator time constants, bias power, heater power range and set point in volts.

The “recorder o/p” panel contains input for the deviation reference of the AVS-47B and a pushbutton “SET” switch. In addition, it has a slide switch selecting the source of the reference: is it obtained from the digital control above or from the LED display.

Note that while the ANALOGUE output of the AVS-47B is always proportional to the sensor resistance, you can affect the DEVIATION output by setting the deviation reference.

The two analog meters for heater power and error signal, familiar from the TS-530A, have been retained. What is extra to the physical controller, is that also the heater current and voltage are calculated and displayed.

Because measuring the heater power etc. is a slow process, we have provided some choices how often this is

made. This is the purpose of the “update” control.

Although multiplexing is possible, its best use is in applications not involving a temperature controller. By shorting the error signal (see proportional gain control) and by latching the integrator (see integrator time constant), you can isolate the controller from the bridge for a while, and measure another sensor. Doing this manually is prone to catastrophic errors, therefore one should build a new application that makes advantage of the SCAN procedure.


Use of the **Application example 1.vi** is almost intuitive to those who have previously worked with the physical AVS-47B and TS-530A. Assume that the control system is ready with a sensor and a heater and a thermal path between them.

After having configured the AVS-47B properly for the control sensor, the reading should stabilise to some reasonable value. Set the reference to a value that requires some heating. The “control error”- meter deflects to positive direction, if more power is needed. Set the “update” to 10 seconds. Select a suitable heater power range. The value depends on situation and can be something from microwatts to one watt. Increase the proportional gain until the power meter has risen a little.


The power meter is updated in every 10 seconds. Once you have found reasonable settings, make the update rate slower, otherwise most of time the system measures the power and not the sensor value at all. As the system approaches equilibrium in the PD mode, enable the integrator with a suitable time constant.

The “History” screen is set for automatic Y-range and the chart history length is 1024 points, dx is 0.1 and the update mode is “strip chart”. You must tailor these

settings according to your own needs with the help of the LabView manual or online reference.



Above is the beginning of the diagram, which is a little complicated. Please open the VI in LabView so that you can better follow this discussion.



After initialisation, the VI forms a big while loop, which is run until the STOP button is pressed. This discussion follows the VISA session “thread”.


At first the program calculates the sum of all AVS-47B front panel controls (figure above). This sum is compared with a hidden control “47sum”. If these are the same, it means that none of the VI front panel controls has been changed, and the **Configure AVS47.vi** can be bypassed. If they are not the same, the program enters a case statement. Inside the case statement, a sum of all AVS-47B controls is calculated again. This sum is different to the above mentioned “47sum”. The new sum is now stored in a local variable, whose name is also “47sum”. When the program exits this case statement, the value in the local variable is stored in the hidden control “47sum”. This is a way to prevent configuring the AVS-47B if none of the settings has changed, and to speed up operation. This **Application example 1.vi** does not use reading initial values from the instruments.

In the next phase, a quite similar case loop configures or does not configure the TS-530A, depending on whether the “530sum” has changed since last comparison.

Of course it is possible to cheat this method by tog-

gling one control up and the other control down so that the sum remains constant. However, it is not easy to do this unintentionally with a mouse in a time less than one cycle.

The deviation reference is set if the SET button is pressed. **Configure reference.vi** needs two inputs, the reference




mode (from reference, from display) and the reference value.

The next small while loop is run only once, because the while-control is always false.

Inside this loop there is one inverter and a *write local variable* “adc”. The inverter gets its value from the *read local variable* “adc” outside the loop. The values stored in these two local variables are Boolean, because they have been created from the ordinary Boolean indicator “adc”. The write local “adc” gets its value inverted from the read local “adc”, and the value in the write local is written to the read local immediately after the case statement has been ended. The effect is that the values oscillate between true and false. The same oscillation is seen in the Boolean indicator “adc” which is the round yellow LED lamp on the left of the display. It changes state on each turn and indicates that the big loop is running, which would otherwise be difficult to see.

One A/D conversion reading is obtained using the **Read one reading.vi**. The integer ADC reading is fed to the “history” chart display. Refer to “Charts” in the LabView manual.

The ADC result is also compared with the TS-530A set point. The difference, which is the control error, is subject to nonlinearising before showing it on the error meter. The error signal is converted to a positive value, 1 is added to prevent it from being zero, and a logarithm is taken. The original sign of the error signal is returned and the result is scaled to suit the meter. The meter is updated



in every turn if “update” is not zero. The purpose of this nonlinearising is to prevent the meter from saturating at high control errors while keeping it reasonably sensitive near to equilibrium.


From the **Read one reading.vi** the VISA session goes into a CASE statement, which serves for calculating the TS-530A heater output. Because measuring the heater output is a very slow procedure, one has to make it only seldom. Otherwise there would remain no time left for measuring the sensor itself, which usually more important than knowing the exact output level.

How often the output is measured is determined by input called “update”. If “update”=0 (“no 530”), the case is identically false. If “update” is not zero, and the current clock time of the computer is equal or greater than the value in the hidden control “interval”, the case will be true and the statement will be run. One operation within the case structure is to add “update” to the current clock time and to store it into a *local write variable* “interval”.

When the program exits the case structure, the value from the local variable is copied into the hidden control “interval”, which will now represent a new moment when the case will become true again.

The heater output voltage and current are read using the **Read from TS530A.vi** two times. The output power is simply the product of these two values. In order to present the power in terms of % of the full range, one has to scale the actual power. This was made by storing coefficients for each range as array elements and by using the “heater range”-control for selecting the element.

One of the most difficult things in writing this VI was to generate the display. For some strange reason, the writers of LabView 5.01 have understood the word “precision” so that it means the number of digits after the decimal point. According to this thinking, 0.01 and 100000.01 would be equally precise.




The A/D converter of the AVS-47B gives 4 1/2 digits, from 0 to 19999. The precision of the reading is limited by the number of significant digits. We wanted the display to indicate these digits, no more or no less regardless of the range setting.

There are 8 possible items to be displayed. First four of them (0..3) can always be expressed in terms of resistance, whereas volts is more suitable for the rest of them. Although the ADC reading could have been used to generate the display with the aid of the range selector, we have used also the floating point R reading, because it takes into account the possible $\Delta R \times 10$ switch.

R is first divided by 1, 1000 or 1E6 so that it is suitable to be expressed either in Ohms, kilohms or Megaohms. The result is led to the display, whose format is determined by the property node. As long as the display item is less than 4, the number of digits after the decimal point (“precision”?) is taken from an array. For display items from 4 up, this number is fixed to 4.

Last, there is a units indicator letter right of the display. This letter is “R”, “k” or “M” for Ohms, kilohms and Megaohms respectively. If the display item is not resistance, the display unit will be “V” for volts.



9.4. Scan2Ch.vi

This is a simple example showing how to start a scan cycle for two sensors and how to read their results.

Autoranging is enabled, and no initial ranges are given. This means that when this VI is used for the first time, the AVS47-IB starts from the 2M Ω range and autoranges as many times as needed to get a proper reading. The correct ranges are remembered, and the next scan cycle should not require this iteration.

Only channels 0 and 1 are measured. Channel 0 is measured using excitation 7 and sensor 1 using excitation 6. Replace the constant with digital controls for any practical application.

When a single scan cycle is started, the status byte output of the AVS47-IB is nulled. The OPC bit will be set once scanning has ended.

The Read Scan Results.vi polls the status until it sees the OPC bit. Then channel 0 is read. In order to read sensor 1, the RdScanCh (=Read Scan Results.vi) is repeated with the timeout control input set to 0. This is the important thing to remember.



INDEX

Symbols

*CLS, clear status command 14
 *ESE, enable standard event command 5, 26
 *IDN? identification query 5
 *OPC, operation complete command 14
 *OPC operation complete command 5
 *RST, reset command 13, 26
 *STB 5
 7210 GPIB controller 13

A

AC excitation 5
 Analog-to-digital conversion, AVS-47 14, 18
 Application example 1.vi 24, 30
 ARN 8
 Autoranging
 disabling 7
 enabled 8, 11
 enabling 7
 in digital filter mode 15
 in scanning 21
 AVE, averaging command 15
 Average count, scanning 12
 Averaging
 average count 15
 filter length 15
 polling STB 13
 within the SCAN procedure 8

B

Bias power, TS-530A 9
 BUF ?, read buffer size and free space 22
 Buffer
 deleting 22
 enabled 7, 16
 in AVS47-IB 10
 Buffer storage 25

C

CASE statement 5
 Channel, AVS47 fp switch 6
 Channel, scanning 11
 Channel number, read scan res. 21
 Channel parameters, scanning 25
 Circular buffer, averaging 15
 Close.vi 6, 13
 compliance voltage, TS-530A 9
 configuration VI's 6
 Configure addresses.vi 5, 8

Configure AVS47.vi 6
 Configure AVS47IB.vi 7
 Configure channel parameters.vi 10, 11
 Configure channel parameters.vi 8
 Configure reference.vi 31
 Configure reference voltage.vi 12
 Configure scan parameters.vi 10
 Configure time and date.vi 10
 Configure TS530A.vi 9
 Continuous scan mode 11, 21
 Control polarity, TS-530A 30

D

Date, configuring 10
 Date, in buffer 22
 Decimal point, AVS-47 display 18
 Default setup.vi 5, 6, 7, 26
 Delay, after stopping 16
 Delay loop
 start single ADC 14
 Derivator time constant, TS-530A 9
 DFR?, read digital filter query 13, 20
 Digital calibration 26
 Digital filter in progress, polling STB 13
 Discuss with AVS47IB.vi 7, 26, 28
 Display, AVS47 fp switch 6, 18
 Downranging 8
 dR magnifier switch, AVS-47 18–27, 20
 DSK, enable buffer storage command 7, 16, 22, 25

E

EEPROM memory 8
 EEPROM write count 8, 25
 Embedded PC 4
 Enable service requests.vi 10
 Enable temperature control ETC, scanning 11
 Enabling buffer storage 25
 Error Message.vi 27
 Error Query.vi 27
 ESB, event status bit in STB 26
 ESR, event status register 5, 26
 Excitation, AVS47 fp switch 6
 Excitation, scanning 12
 Excitation frequency 5

F

File name 22
 Filter length
 averaging 15
 Firmware version 5, 28
 First channel FCH, scanning 11
 Front panel started measurements 7
 FSM 5, 7
 FSM, front panel started measurements 25



FSM enabled, AVS47IB 7

G

Getting started.vi 23, 29

Go local.vi 6, 13

Go remote.vi 12, 29

GPIB

device address 5, 8, 25, 28, 29, 30

readdressing 28

H

HDR 5

HDR, message headers 25

Heater, TS-530A

current 23

heater resistance 9

power 23, 32

power range 9

voltage 23

Heater current, TS-530A 23

Heater power, TS-530A 23

Heater voltage, TS-530A 23

Hidden controls, in LabView 31

I

IBA, GPIB bus address command 25

Identification query 5

Idle state, polling STB 13, 17

IEEE-488.2

status byte 5, 13

status reporting 26

Initialize.vi 4, 5, 7, 12, 26, 29

Initial values, generating by reading 25, 29, 31

Input, AVS47 fp switch 6

Instrument front panels 28

Integration time constant, TS-530A 9

K

KIL, delete buffer file command 22

L

LabView

driver standards 4–27

version number 4

VI defaults 6, 7, 8

Last channel LCH, scanning 11

Local control mode 4

Local variables, in LabView 29, 31

Loop delay

in WaitFor 15, 19, 21

LPT1: 7

M

MAV message available 5, 10, 13, 15, 18, 26

Maximum, reading average 19

Message headers 7

Message headers included, AVS47IB 7

Minimum, reading average 19

Multiplexing 8

Multitasking 5

N

National Instruments

GPIB-PCII card 4

LabView driver standards 4, 5, 26, 27

Not in idle state

error indicator 7

NXT?, get next buffer line query 14, 22

O

Omitted VI' 27

Operation complete event 5, 15, 18, 26

Operation complete OPC 10, 13, 24

Overrange 7, 11, 19

detecting in scanning 21

OVL ?, overrange query 19

P

Parallel printer 7

PBA 8

PBA, Picobus address 25

PBD, Picobus delay factor command 25

PD control mode 9

Picobus

address 8

cable 18

delay factor PBD 8

speed 8

transaction 22

Picolink

Optical fibre link 18

PID parameters 9

reading from TS-530A 24

Poll STB status.vi 7, 10, 12, 13, 14, 15, 20

power-on

defaults AVS47-IB 5, 7, 8, 11

power bias, TS-530A 9

Precision, LabView interpretation 32

Printer enabled, AVS47IB 7, 25

printer port of the AVS47-IB 7

Printing the buffer on printer, polling STB 14

PRN 7, 25

Programmatical use of the VIs 30

Programming mode, config. ref. volt 12

Property node

VISA open 28

proportional gain, TS-530A 9



R

RAM disk buffer 7
 Range, AVS47 fp switch 6
 Range, scanning 12
 RBF, read buffer command 22
 Read 47IB parameters.vi 25
 Read and save buffer.vi 7, 17, 21
 Read average.vi 19, 21
 Read AVS47 parameters.vi 22, 29
 Read channel parameters.vi 25
 Read digital filter.vi 15, 20
 Read from TS530A.vi 23, 32
 Reading the buffer in RBF mode, polling STB 14
 Read one reading.vi 18, 31
 Read scan parameters.vi 25
 Read scan results.vi 17, 21
 Read single ADC.vi 17, 18
 Read TS530A parameters.vi 24
 Real-time data acquisition 4
 Real time clock, AVS47-IB 10
 Reference voltage, config ref. volt. 12
REM
 disabling remote control 6, 13
 enabling 12
 remote control command 5, 6
 Remote control mode 4, 12
 RES?, read resistance query 18
 Reset 5
 state of AVS47, TS530A 5
 Reset.vi 26
 Resource name, VISA 28
 Response message headers 5, 7
 Revision Query.vi 27
 RF filters 8
 Ring controls 6
 Running average 15

S

Scan2Ch.vi 33
 Scan interval delay, polling STB 13
 Scan interval SCI 11
 Scanning 8
 Scanning, polling STB 13
 SCAN procedure 8, 10, 11, 15, 20, 25, 30
 SCN, start scan cycle command 17
 Self-Calibration in progress, polling STB 14
 Self-test is in progress, polling STB 14
 Self calibrate.vi 26
 Self test.vi 26
 Serial polling 5, 10, 13, 16
 Service request, GPIB 10
 Set point, TS-530A 23
 set point. TS-530A 9
 Set point DAC, TS-530A 23

Single scan cycle 21
 spreadsheet form, save buffer 22
 SRE service request enable register 10
 SRQ service request 13
 Stabilisation delay 8
 Stabilisation delay, AVS47IB 8
 Stabilisation delay, scanning 11
 Standard deviation, reading average 19
START/PRINT
 AVS-47 front panel switch 5, 7
 Start averaging.vi 15
 Start digital filter.vi 15
 Start single ADC.vi 14, 17, 18
 Start single scan.vi 16, 21
 State of the instruments 4
 Status reporting 10
 STB status byte 5, 13
 Steady-state error 9
 Stop.vi 15, 16
 STP, stop command 13

T

Take one reading.vi 29
 Temperature control period TCP, scanning 11, 13
 Temporary temperature control , polling STB 13
 Time, configuring 10
 Time, in buffer 22
 Timeout 15, 17, 19
 in Discuss.vi 26
 in reading average 19
 in scanning 21
 TST, self test query 28

U

Upranging 8

V

Version of the firmware 5
VISA
 read 5, 19, 22
 read STB 5
 session 5, 6, 28, 32
 Virtual Instrument Standard Architecture 4
 write 23

W

Wait for OPC/MAV/TMO.vi 15, 17, 19, 21, 23, 29
 Waiting for the START/PRINT, polling STB 13
 Wait SRQ, VISA function 10
 Windows 3.11 4



10. Revision History

1R0 => 1R1

Effect of the “system decimal point” on reading operations from the AVS47-IB was removed by forcing LabView to use period regardless of the setting in Windows.

The Wait for OPC/MAV/TMO.vi was modified so that if the timeout input control is set to zero, no timeout error output is generated.

A simple example on scanning two sensors “Scan2Ch.vi” was included in the driver set of files.

NEW PRINTOUT

This manual was converted to InDesign format 2006-10-04. Only small error corrections.

1R1 => 1R2

Major revision, where the driver was compiled with LabView version 7.1 (driver 1R1 was compiled with LV 5.01). Some VIs were changed and some bugs corrected.